

How long it usually takes for the liquid level in the manometer tube of a fisher sub sieve analyzer reaches a maximum?

--It usually takes at least 5 min for the liquid level rise slowly until it reaches a maximum during fisher number particle sizing.