

Apparatus of Hall Flow Meter according to ISO 4490 about apparent bulk density?

-The funnel is made of a non-magnetic, corrosion-resistant metallic material having sufficient wall thickness and hardness to withstand distortion and excessive wear.

[AS-300 Hall Flow Meter](#)